

OFFICE OF THE PURCHASE COMMITTEE NO:-II  
(RATE CONTRACT COMMITTEE)  
HEALTH & MEDICAL EDUCATION DEPARTMENT  
**GOVT. MEDICAL COLLEGE JAMMU.**

-\*-

**Sub:- Additions/Alterations in Tendered specifications of Machinery & Equipment for the year 2010-11**

As per discussions in Pre Bid Meeting on tendered specifications of Machinery & Equipments for the year 2010-11 (Additional items) held on 06-12-2010 in the office chamber of Director Health Services Jammu and representations received thereof from the different firms, alterations/deletion/ additions made by the concerned Head of the departments are mentioned below against each item for information of the participating tenderers. Rest of the specifications shall remain same:-

Item No:	Please Read
01	<p><b>A) Pump (Quaternary).</b> <u>Pressure range:</u> 0 to 60 MPa (600 bar, <b>6000 Psi</b>) upto 5ml/min</p> <p><b>C) Auto Sampler:</b> <u>Precision typically</u> &lt; 0.3% RSD of peak areas from &lt;5µL to 100µL .</p> <p><b>E) Photo Diode Array Detector</b> <b>a) <u>Wavelength range:</u></b> 190-800nm or more, <b>g) <u>Detector types:</u></b> 512 elements Diode Array. <b>k) <u>Thermostated</u> Standard flow cell of 10mm.</b></p>

**Sd/-**  
**(Member Secretary)**  
Purchase Committee No-II  
Health & Medical Education Department  
Govt. Medical College Jammu

No: GMC/PC-II/ 6376-85

Dated:-16-12-2010

Copy to the:-

1. Principal, Govt. Medical College Jammu / Srinagar for information
2. Director Health Services Jammu / Srinagar for information
3. Vice Chairman Purchase Committee No: -II, (HOD ENT) Govt. Medical College Jammu for information.
4. Joint Director, Information Department Jammu for information and with the request to get the notice published in atleast two leading local papers with largest circulation for wider publicity and cutting be sent to this office for confirmation.
5. All concerned for information.